

Application of the Mesh Experiment for the Back-Illuminated CCD: I. Experiment and the Charge Cloud Shape

Emi MIYATA, Masami MIKI, Junko HIRAGA, Hirohiko KOUNO, Kazutoshi YASUI, Hiroshi TSUNEMI, Kazuhisa MIYAGUCHI², and Kouei YAMAMOTO²

*Department of Earth & Space Science, Graduate School of Science, Osaka University,
1-1 Machikaneyama, Toyonaka, Osaka 560-0043, Japan*

¹ *Solid State Division, Hamamatsu Photonics K.K.,
1126-1 Ichino-cho, Hamamatsu City 435-8558, Japan*

(Received)

We have employed a mesh experiment for back-illuminated (BI) CCDs. BI CCDs possess the same structure to those of FI CCDs. Since X-ray photons enter from the back surface of the CCD, a primary charge cloud is formed far from the electrodes. The primary charge cloud expands through diffusion process until it reaches the potential well that is just below the electrodes. Therefore, the diffusion time for the charge cloud produced is longer than that in the FI CCD, resulting a larger charge cloud shape expected.

The mesh experiment enables us to specify the X-ray point of interaction with a subpixel resolution. We then have measured a charge cloud shape produced in the BI CCD. We found that there are two components of the charge cloud shape having different size: a narrow component and a broad component. The size of the narrow component is $2.8 - 5.7 \mu\text{m}$ in unit of a standard deviation and strongly depends on the attenuation length in Si of incident X-rays. The shorter the attenuation length of X-rays is, the larger the charge cloud becomes. This result is qualitatively consistent with a diffusion model inside the CCD. On the other hand, the size of the broad component is roughly constant of $\simeq 13 \mu\text{m}$ and does not depend on X-ray energies. Judging from the design value of the CCD and the fraction of each component, we conclude that the narrow component is originated in the depletion region whereas the broad component is in the field-free region.

KEYWORDS: charge-coupled device, X-ray event, split event, subpixel spatial resolution, mesh experiment

1. Introduction

A charge-coupled device (CCD) is widely used in an optical imaging as well as an X-ray imaging. When an X-ray photon is photoabsorbed in the CCD, it generates a primary charge cloud consisting of thousands of electrons. The size of the primary charge cloud is $73 \times (E/2.3 \text{ keV})^{1.75} \text{ nm}$ where E is energy of X-ray photon in unit of keV.¹⁾ The primary charge cloud expands through diffusion process until they reach the potential well of the CCD pixel. The final charge cloud after diffusion process is collected into several pixels forming various types of event pattern depending on how they split.

We have developed a new technique “mesh experiment” which enables us to restrict the X-ray point of interaction with a subpixel resolution.²⁾ Hiraga *et al.*³⁾ investigated the event pattern produced and measured the charge cloud shape produced in the front-illuminated (FI) CCD for the first time. They found that a charge cloud shape could be well represented by a Gaussian function. They also obtained the standard deviation, σ , of the final charge cloud to be $0.7 \sim 1.5 \mu\text{m}$ for $1.5 \sim 4.5 \text{ keV}$ X-rays by using a FI CCD. Based on this experiment, they confirmed that there are three parameters tightly coupled together;⁴⁾ an X-ray point of interaction within a pixel, an X-ray event pattern and a charge cloud shape. Any two parameters can determine the third one. The event pattern is quite easily noticed while the charge cloud shape is difficult to be measured. Currently, they can separately measure it only by using the mesh experiment. Therefore, they can determine the X-ray point of interaction with much better spatial resolution than the pixel size. They obtained the position resolution of $0.7 \mu\text{m}$ using the pixel size of $12 \mu\text{m}$ if an X-ray photon became a split pixel event. However, due to its relatively small size of the charge cloud, the fraction of split pixel events whose point of interaction can be improved is less than 10 % of the total events.

In this paper, we applied the mesh experiment to the back-illuminated (BI) CCD. Figure 1 shows the cross-section and the potential profile of FI CCDs. In the case of FI CCDs, since X-ray photons enter from the front side where there are electrodes. A major part of X-rays are photo-absorbed close to electrodes, resulting a final charge cloud to be relatively small. Figure 2 shows the cross-section and the potential profile of BI

CCDs. In the case of BI CCDs, a substrate of CCD, shown in Fig 1 (a), is etched from the back toward the depletion layer (thinning process) in order to achieve high quantum efficiency for blue lights and low energy X-rays. After the thinning process, almost all of the substrate has been removed. The thickness of the BI CCD employed is $\simeq 20\text{ }\mu\text{m}$. Since X-ray photons enter from the back surface of the CCD shown in Fig 2, a primary charge cloud is formed far from the electrodes. The primary charge cloud expands through diffusion process until they reach the potential well which is just below the electrodes. Therefore, the diffusion time for the charge cloud generated by X-rays is longer than those in the FI CCD, resulting that a larger charge cloud shape is expected.

Low energy X-rays, however, are photo-absorbed in the field-free region of BI CCDs. The primary charge cloud expands due to a diffusion process toward the electrodes as well as the back surface. The charge traveled to the back surface will be lost due to a recombination process. In the production of BI CCDs, accumulation is the most significant process that gives an internal potential to the back surface of CCDs and repels the signal charge to the electrodes.⁵⁾ In the case of the BI CCD employed, the accumulation process is performed by ion implantation.⁶⁾ As shown in Fig 2 (b), the potential at close to the back surface is locally lower than that at the inner region caused by the accumulation process which enables us to collect the charge generated close the back surface.

2. Experiment and Results

The basic idea and the algorithm of data reconstruction for the mesh experiment are given in Tsunemi *et al.*²⁾ We performed mesh experiment employing a BI CCD (S7170) fabricated by Hamamatsu Photonics K.K. The BI CCD has a two-phase electrode and full-frame transfer type. It possesses 512×512 active pixels with a size of $24\text{ }\mu\text{m}$ square. The mesh made of gold has small holes of $2.1\text{ }\mu\text{m}$ diameter with spaced $48\text{ }\mu\text{m}$ apart. The pitch of mesh holes is twice larger than that of the CCD pixel. The mesh is placed just 5 mm above the CCD. We employed the 21 m long X-ray beam line in our laboratory. We used the X-ray generator of Ultra-X18, fabricated by RIGAKU, with Al, Mo, and Ti targets in order to irradiate the characteristic emission line of each target: Al-K (1.5 keV), Mo-L (2.3 keV), and Ti-K (4.5 keV). We applied the voltage of 6, 5, and 7 kV for Al, Mo,

and Ti target, respectively. We used a mechanical shutter made of stainless steel having thickness of $60\mu\text{m}$ to control the X-ray beam so that the pile-up would not become a serious effect. We drove the BI CCD with our newly-developed system named *E-NA* system.⁷⁾ The CCD analog data were processed by an integration-type circuit⁸⁾ and digital data were acquired with a VME system. We controlled the operating temperature of the CCD to be -100°C during the experiment.

Figure 3 shows X-ray spectra of single-pixel events obtained with three targets employed. The readout noise of our system is $\simeq 20\text{e}^-$ -rms. Each spectrum shows the characteristic emission line superimposed on the continuum emission. We should note that the peak channel of Mo-L emission line is close to that of Al-K emission line whereas the energy is quite different.

We selected events between two dashed lines shown in Fig 3 in the subsequent analysis.

2.1 Event pattern

Figure 4 shows a part of a raw image obtained with single-pixel events. We can see a clear moiré pattern. Based on the moiré pattern obtained, we can determine the geometrical relation between the mesh and the CCD. Details of the method are given in Tsunemi *et al.*⁹⁾

Once we find their relations, we can determine the X-ray point of interaction within a pixel for all X-ray events. We thus rearranged each X-ray event with a subpixel resolution and investigated the distribution of single-pixel event, two-pixel split event, three- and four-pixel split event and more extended events as a function of the point of interaction inside a pixel. Figure 5 shows their X-ray intensity maps in 3×3 representative pixels (RP) for Mo-L X-rays. Each black dot corresponds to an individual X-ray event. When X-ray photons enter the center of a pixel, a major part of them produce single-pixel events shown in (a). When the X-ray point of interaction becomes close to the horizontal or vertical boundary of each pixel, the horizontally or vertically two-pixel split events are formed shown in (b) or (c). When an X-ray photon enters the corner of each pixel, it produces three- or four-pixel split event shown in (d). This tendency is consistent with that obtained for the FI CCD.¹⁰⁾ However, we found that the regions forming the three-

and four-pixel split events become larger and extend into inner region than FI results (e.g. Figure 7 of Tsunemi *et al.*⁹⁾) whereas the pixel size of the BI CCD employed is $24\mu\text{m}$ that is two times larger than that of the FI CCD. This suggests that a charge cloud size produced in the BI CCD is much larger than that of the FI CCD. This fact is clarified in Fig 5 (e) since a charge cloud spreads into more than four pixels even if an X-ray enters the center of a pixel.

2.2 Absorption structure of BI CCDs

As shown in Fig 1, X-ray photons enter from the electrodes in the case of FI CCDs. The thickness of electrodes is not uniform and some part of electrodes is overlapped, resulting the detection efficiency within a pixel to be far from uniform.^{2, 11–13)} It makes it difficult to make the response matrix of FI CCDs.

As to the BI CCDs, on the other hand, X-ray photons enter from the back surface where the thickness of the field-free region is much thinner than those of FI CCDs shown in Fig 2. The potential created by the accumulation process also drives the charge generated in the field-free region toward the electrodes, resulting higher detection efficiency for low energy X-rays. Since the thickness of the back surface of CCDs is uniform, we can expect a uniform detection efficiencies within pixels for BI CCDs. Figure 6 shows the X-ray intensity map extracted from all X-ray events in 3×3 RPs for Mo-L X-rays. There is no significant variation in each image, demonstrating the uniform detection efficiency of the BI CCD. We also confirmed uniformity for Al-K and Ti-K X-rays. The non-uniformities in detection efficiency are $\simeq 2.5$, 2.0 , and 7.2% in a standard deviation for Al-K, Mo-L, and Ti-K X-rays, respectively. This uniformity is a large advantageous point of BI CCDs over FI CCDs. We should note that the uniformity of Ti-K X-rays is lower than those of other X-rays. It must be due to the fact that the mean attenuation length of Ti-K X-rays is long enough for X-ray point of interaction to be affected by electrodes.

2.3 Amount of charge collected in a center pixel

When X-rays enter the CCD at the position of (X_p, Y_p) , the output of the n_{th} pixel of the CCD, $P_n(X_p, Y_p)$, is written as

$$P_n(X_p, Y_p) = \int_{X_n}^{X_{n+1}} dX \int_{Y_n}^{Y_{n+1}} dY f(X - X_p, Y - Y_p)$$

where $(X_n, X_{n+1}, Y_n, Y_{n+1})$ denotes the boundary of the n_{th} pixel and $f(X, Y)$ is the charge cloud shape just before the collection by the potential well. The actual output of n_{th} pixel, $D_n(X_{in}, Y_{in})$, is a convolution between $P_n(X_p, Y_p)$ and the mesh hole, which can be described by

$$D_n(X_{in}, Y_{in}) = \int dX_p \int dY_p P_n(X_p, Y_p) H(X_{in} - X_p, Y_{in} - Y_p) \quad (2.1)$$

$$= \int_{X_n}^{X_{n+1}} dX \int_{Y_n}^{Y_{n+1}} dY f \otimes H(X - X_{in}, Y - Y_{in}) \quad (2.2)$$

where (X_{in}, Y_{in}) is the center of mesh hole, $H(X, Y)$ is a typical hole shape of the mesh, and $f \otimes H$ represents the convolution between f and H .

We can experimentally measure $D_n(X_{in}, Y_{in})$ by using the RP generated with all X-ray events. Figure 7 show the surface profiles and projected profiles of D_n in 3×3 pixel regions where the center pixel is n_{th} pixel for Mo-L, Al-K, and Ti-K X-rays.

Hiraga *et al.*³⁾ measured D_n for the FI CCD for the first time and found a flat-top feature around the center of n_{th} pixel. This is due to the fact that the charge collected in the n_{th} pixel is almost constant, forming single-pixel events, when a mesh hole is well within the pixel. D_n obtained with Ti-K for the BI CCD has relatively flat top at the center of pixel. On the other hand, D_n shows a sharp structure even near the center of pixel in Mo-L and Al-K X-rays, suggesting that a charge splits into neighboring pixels even if X-rays enter near the center of the pixel. We can again expect that a charge cloud size produced in the BI CCD is much larger than that of the FI CCD. It is interesting to note that a flat-top feature obtained only with Ti-K X-rays suggests that a size of the charge cloud of Ti-K X-rays is smaller than those obtained with Mo-L and Al-K, which is in contrast to those obtained with FI CCDs.¹⁰⁾

2.4 Charge cloud shape

Hiraga *et al.*³⁾ calculated a charge cloud shape by differentiating D_n (we hereafter referred to the differential method). As shown in Fig 7, however, we cannot obtain a charge cloud shape with the differential method since D_n does not show a flat-top feature for the BI CCD. We then employed a different approach to obtain a charge cloud shape.

As written in eq 2.1, D_n is an integration of convolution between a charge cloud shape and a mesh hole within a CCD pixel. We directly calculate their convolution to reproduce D_n for each energy assuming a charge cloud shape to be a two-dimensional axial-symmetric Gaussian function³⁾ (the integral method, hereafter). We should note that we have applied the integral method to the FI data and obtained consistent results to those obtained by the differential method. We found, however, a difficulty to reproduce D_n with the integral method. The problem is that our method cannot reproduce a sharp component and a tail component, simultaneously. Such a tail component is detected for all energies whereas it was not detected with FI CCD.³⁾ The presence of an tail component suggests a large size of a charge cloud which must be much larger than that forming a sharp component. As shown in Fig 5 (e), the events spreading more than 4 pixel is uniformly distributed within a pixel. These events should account for such tail components.

We thus introduce an extra component to fit the data. We employed two components in the charge cloud shape having different size in both axes: a narrow component for the sharp structure and a broad component for the tail structure. In this way, we found that the two-component model in the integral method well reproduced D_n obtained. The model functions of D_n are shown in Fig 8. Table I summarizes the charge cloud sizes of two components for all energies. Figure 9 shows the size of the charge cloud for both components as a function of the X-ray attenuation length in Si.

The broad component of the charge cloud is 2.5 – 5 times larger than that of the narrow component. The size of the narrow component decreases as the attenuation length increases whereas that of the broad component is relatively constant of $\sim 13 \mu\text{m}$. Comparing the FI results,¹⁰⁾ the sizes for the narrow and the broad component are 2–8 times larger and an order of magnitude larger, respectively. The number of X-ray events of the broad component is ~ 3 times larger than that of the narrow component.

3. Discussion

3.1 *Non-linearity of pulse height versus energy*

As shown in Fig 3, the pulse height of each characteristic X-ray line does not linearly increase with energy. Figure 10 shows their relationship between the pulse height and the incident X-ray energy. The pulse height of Mo-L is $\simeq 30\%$ lower than that expected from the linear relationship based on Al-K and Ti-K X-rays as shown by a solid line. This is quite a different characteristics of the BI CCD to the FI CCD.

The attenuation length of Mo-L X-rays is only $2.2\ \mu\text{m}$ that is the shortest among X-rays employed. Therefore, most of Mo-L X-rays are absorbed close to the accumulation region, where the density of impurities is an order of magnitude larger than that in the depletion region, leading to the recombination between impurities and signal charge. We thus suppose that the non-linearity of pulse height versus energy is caused by an incomplete collection of the signal charge produced in the accumulation or some part of the field-free region.

3.2 *Origin of the two components of charge cloud*

We found that there were two components in a charge cloud shape generated by the BI CCD. A narrow components has a σ of $2.8 - 5.7\ \mu\text{m}$ which strongly depends on the incident X-ray energy. The size of a broad component is relatively constant of $\simeq 13\ \mu\text{m}$ having less dependence on the X-ray energy whereas such a component is not appeared in the FI CCD.³⁾

As to BI CCDs, X-rays entering from the back surface of the CCD form a primary charge cloud far from the electrodes. The primary charge cloud expands through diffusion process until they reach the potential well which is just below the electrodes. Therefore, a diffusion time for the charge cloud produced by low energy X-rays is longer than that by high energy X-rays, resulting a larger size of charge cloud shape for low energy X-rays. This relation is consistent with that obtained for a narrow component. However, we found that the size of the broad component is almost independent of the incident X-ray energy.

As shown in Figs 1 (b) and 2 (b), the potential profiles of FI and BI CCDs are different in the field-free region. A charge generated in the field-free region can be collected only

through the diffusion process in FI CCDs. On the other hand, the potential created by the accumulation drives the charge generated in the field-free region toward electrodes in the BI CCD. Therefore, a charge generated both in the depletion region and in the field-free region can be collected in the BI CCD. With taking into account the difference in the charge cloud size for both components, we suppose that an X-ray photo-absorbed in the depletion region can be detected as a narrow component whereas an X-ray photo-absorbed in the field-free region can be detected as a broad component.

Miyata *et al.*⁸⁾ measured the thickness of the depletion depth of FI CCD having similar wafer as that of the BI CCD employed to be $\simeq 4\ \mu\text{m}$. Since the thickness of the BI CCD employed is $20\ \mu\text{m}$, thickness of the field-free region is $\simeq 16\ \mu\text{m}$. This value is in good agreement with the charge cloud size of the broad component if the charge cloud expands spherically in the field-free region. Moreover, as shown in Table I, the fraction of the broad component obtained is ~ 3 times larger than that of the narrow component, which is also consistent with the ratio of thickness between the field-free region and the depletion region. These facts strongly support that a broad component is originated in the field-free region.

3.3 Asymmetry of the charge cloud shape

The charge cloud size obtained with the BI CCD is shown in Table I. Both for the narrow and broad components, the charge cloud shapes are almost point symmetric. The difference in both directions is less than 5%. This is also noticed that D_n obtained for three energies shown in Fig 7 show a quite point-symmetric shape.

Hiraga *et al.*³⁾ pointed out that a charge cloud shape obtained for the FI CCD shows an asymmetric shape. The size in unit of the standard deviation for the x-direction was $1.4\ \mu\text{m}$ whereas that for the y-direction was $0.7 - 0.8\ \mu\text{m}$. Tsunemi *et al.*¹⁴⁾ further investigated the possible explanations about the asymmetry of the charge cloud shape. They conclude that the asymmetry is due not to the experimental setup, such as the asymmetric mesh hole, but to the asymmetry of the electric field inside the CCD. Similar asymmetry has obtained for some other FI CCDs.^{10,11)} Therefore, the asymmetric shape of the charge cloud is commonly found in FI CCDs.

On the other hand, we obtained a symmetric shape of the charge cloud generated inside the BI CCD for all three X-rays. The back surface of CCDs are uniform and is normally applied to be zero voltage. At the front side, there are four electrodes per pixel in our case and they are slightly overlapped to each other as shown in Fig 1 (a). Moreover, at the front side, there is another gate structure, “channel stop”, which forms the pixel boundary along the y-direction. All these gate structures generate the electric field not to be uniform near the front electrodes as pointed out by Tsunemi *et al.*¹⁴⁾ The symmetry of the charge cloud obtained with BI CCD suggests that a charge cloud must not be affected by the electrodes significantly because it is generated far from the electrodes.

4. Conclusion

We have performed the mesh experiment with the BI CCD and measured the charge cloud size for the first time. We found that there are two components of the charge cloud shape to reproduce the data obtained all for Mo-L, Al-K and Ti X-rays: a narrow component and a broad component. The narrow component possesses a σ of $2.8 - 5.7 \mu\text{m}$ which strongly depends on the attenuation lengths of incident X-ray energies. The broad component, on the other hand, possesses much larger σ than that of the narrow component. It is $\simeq 13 \mu\text{m}$ and almost independent of the attenuation length of X-rays. The accumulation technique to fabricate BI CCDs is not employed in the FI CCDs and the charge produced even in the field-free region can be collected toward electrodes. We thus suppose that the narrow and the broad components are originated in the depletion region and the field-free region, respectively. This hypothesis is consistent with both by the design value of the CCD and by the fraction of X-ray events for both components.

We have found that the linearity between the incident X-ray energies and the pulse height is not good. The pulse height of Mo-L X-rays is $\simeq 30\%$ lower than that expected. This might be explained by the fact that the effect of the recombination with the impurities is significant in the accumulation region since the attenuation length of Mo-L is shortest among those X-rays employed.

Acknowledgement

J.H. is partially supported by JSPS Research Fellowship for Young Scientists, Japan. This work is partly supported by the Grant-in-Aid for Scientific Research by the Ministry of Education, Culture, Sports, Science and Technology of Japan (13874032, 13440062).

References

- 1) T.E. Everhart and P.H. Hoff: J. Appl. Phys. **42** (1971) 5837.
- 2) H. Tsunemi, K. Yoshita and S. Kitamoto: Jpn. J. Appl. Phys. **36** (1997) 2906.
- 3) J. Hiraga, H. Tsunemi, K. Yoshita, E. Miyata and M. Ohtani: Jpn. J. Appl. Phys. **37** (1998) 4627.
- 4) H. Tsunemi, J. Hiraga and E. Miyata: Nucl. Instrum. and Method, **477** (2002) 155.
- 5) J. Janesick, T. Elliott, T. Daud and J. McCarthy: Proc SPIE **570** (1985) 46.
- 6) M. Muramatsu, H. Akahori, K. Shibayama, S. Nakamura and K. Yamamoto: Proc SPIE **3019** (1997) 2.
- 7) E. Miyata, C. Natsukari, D. Akutsu, T. Kamazuka, M. Nomachi and M. Ozaki: Nucl. Instrum. and Method, **459** (2001) 157.
- 8) E. Miyata, C. Natsukari, T. Kamazuka, H. Kouno, H. Tsunemi, M. Matsuoka, H. Tomida, S. Ueno, K. Hamaguchi and I. Tanaka: Nucl. Instrum. and Method, (2002) in press.
- 9) H. Tsunemi, J. Hiraga, K. Yoshita and S. Kitamoto: Jpn. J. Appl. Phys. **37** (1998) 2734 .
- 10) H. Tsunemi, J. Hiraga and E. Miyata: Nucl. Instrum. and Method, **436** (1999) 32.
- 11) K. Yoshita, H. Tsunemi, E. Miyata, K.C. Gendreau and M.W. Bautz: IEEE Trans. Electron Devices, **46** (1999) 100.
- 12) M. J. Pivovarovoff, S. Jones, M. W. Bautz, S. Kissel, G. Prigozhin, G. Ricker, H. Tsunemi and E. Miyata: IEEE Trans. NS, **45** (1998) 164.
- 13) J. Hiraga, H. Tsunemi, A.D. Short, A.F. Abbey, P.J. Bennie and M.J.L. Turner: Nucl. Instrum. and Method, **465** (1999) 384.
- 14) H. Tsunemi, J. Hiraga, K. Yoshita and E. Miyata: Jpn. J. Appl. Phys. **38** (1999) 2953.

Figure captions

Fig. 1. (a) Cross-section and (b) potential profile of a FI CCD.

Fig. 2. Same as Fig 1 but for a BI CCD.

Fig. 3. X-ray spectra obtained with targets of (a) Al, (b) Mo, and (c) Ti. Each characteristic emission lines are identified. The dotted lines show the ranges employed for analysis.

Fig. 4. A part of the raw image (200×200 pixels) extracted from the single-pixel events of Mo-L X-rays. Each dot shows an X-ray event. A clear Moire pattern can be seen. Since the pitch of the mesh hole is two times that of the CCD pixel, we find X-ray events appeared in every 2×2 pixels.

Fig. 5. X-ray intensity map of (a) single-pixel events, (b) horizontally two-pixel split events, (c) vertically two-pixel split events in 3×3 RPs, (d) three- and four-pixel split events, and (e) more extended events with the subpixel resolution. The dotted lines show the pixel boundary.

Fig. 6. X-ray intensity map extracted from all X-ray events in 3×3 RPs for Mo-L. The uniform detection efficiency can be found for BI CCD. Projected profiles onto X- and Y-axes are also shown. The dotted lines show the pixel boundary.

Fig. 7. Amount of charge collected in the n_{th} pixel, D_n , is shown in 3×3 pixels for (a) Mo-L, (b) Al-K, and (c) Ti-K X-rays. Upper panel of each figure shows the 3-d picture of D_n and projected profiles are shown in lower panel.

Fig. 8. Model of D_n for (a) Mo-L, (b) Al-K, and (c) Ti-K X-rays.

Fig. 9. Charge cloud size as a function of the X-ray attenuation length in Si for (a) the narrow component and (b) the broad component.

Fig. 10. Relationship between the pulse height and incident X-ray energy. The solid line shows the linear relationship between them to reproduce Ti-K X-rays.

Table I. Charge cloud shape obtained with the BI CCD for Mo-L, Al-K, and Ti-K X-rays.

	Mo-L	Al-K	Ti-K
Energy [keV]	2.3	1.5	4.5
Attenuation length [μm]	2.2	7.9	13.5
Narrow component			
σ_{nx}	5.3 ± 0.3	4.3 ± 0.2	2.91 ± 0.08
σ_{ny}	5.8 ± 0.3	4.4 ± 0.2	2.99 ± 0.09
S_{nx}	5.3 ± 0.3	4.2 ± 0.1	2.83 ± 0.08
S_{ny}	5.7 ± 0.3	4.3 ± 0.2	2.91 ± 0.09
Broad component			
σ_{wx}	13.4 ± 0.3	13.2 ± 0.3	11.9 ± 0.1
σ_{wy}	13.9 ± 0.3	13.8 ± 0.3	12.5 ± 0.1
S_{wx}	13.4 ± 0.3	13.2 ± 0.3	11.9 ± 0.1
S_{wy}	13.9 ± 0.3	13.8 ± 0.3	12.5 ± 0.1
Fraction of narrow component			
	0.25 ± 0.02	0.32 ± 0.01	0.32 ± 0.01

This figure "f1.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f2.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f3a.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f3b.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f3c.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f4.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f5a.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f5b.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f5c.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f5d.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f5e.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f6.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f7.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f8.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f9a.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f9b.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>

This figure "f10.jpg" is available in "jpg" format from:

<http://arXiv.org/ps/physics/0206021v1>